



## SPECIFICATIONS

Part No.	Alloy	Adapter Style	Center Hole	Product Weight	Build Height	Max Patient Weight	Set Screw Torque (M8x12mm)	Clamp Screw Torque (M5x16mm)
3T18	Titanium	Pyramid	8.2 mm	71 gr.	17 mm.	136 Kg. / 300 lbs.		9 Nm
3T19	Titanium	Pyramid	10 mm	71 gr.	17 mm.	136 Kg. / 300 lbs.		9 Nm
3T23	Titanium	Receiver		80 gr.	21 mm.	136 Kg. / 300 lbs.	15 Nm.	9 Nm
3S18	S. Steel	Pyramid	8.2 mm	124 gr.	17 mm.	136 Kg. / 300 lbs.		9 Nm
3S23	S. Steel	Receiver		130 gr.	21 mm.	136 Kg. / 300 lbs.	15 Nm.	9 Nm
3A18	Aluminum	Pyramid	8.2 mm	99 gr.	17 mm.	100 Kg. / 220 lbs.		9 Nm
3A23	Aluminum	Receiver		74 gr.	21 mm.	100 Kg. / 220 lbs.	15 Nm.	9 Nm

## USAGE

This product is designed to be used for lower limb external prosthetic applications. Loosen clamp screw and rotate to desired position. Before fitting this product onto the patient, all parts must be joined properly and all set screws must be secured enough for the patient to walk. In order to make final adjustments, there should be no loose screws.

## ATTENTION

After the final adjustments, apply single drop of removable Loctite® 242 Blue® to each of the set/cap screws and then torque the product. Improper installation or uneven, insufficient or excessive torque may cause bolt or adapter failure. Always use quality metric tools for installation.

## WARRANTY

All MEDEX Modular Adapters are unconditionally guaranteed for 2 years. Cutting, drilling, modifying or using the component other than as described in this manual will void the warranty.